

## LV7390 Software Upgrade Procedures

**Caution: Please DO NOT turn off the power during software is upgraded.**

1. Unzip the **xxxx.zip** file.
2. Copy the "**LV7390\_UPDATE**" folder to the USB thumb drive from the PC.  
(The LV7390\_UPDATE folder includes only one file)  
NOTE: The USB thumb drive maximum capacity should be 4GB and it should be formatted as FAT32.
3. Plug in the USB thumb drive to the LV7390.
4. Power ON the LV7390.
5. Press the **SYS** button on the front panel.
6. Press **[F7] INITIALIZE**
7. Press **[F7] FIRMWARE UPDATE**
8. Press **[F3] UPDATE YES**
9. Update will take a few minutes and it will reboot the instrument after update is completed.
10. END